

Secretary of State
Certificate and Order for Filing
TEMPORARY ADMINISTRATIVE RULES
Generated on June 24, 2013 10:47AM
DRAFT

Oregon Board of Dentistry	818
Agency and Division	Administrative Rules Chapter Number
Stephen Prisby	stephen.prisby@state.or.us
Rules Coordinator	Email Address
1600 SW 4th Ave., Suite 770, Portland, OR 97201	971-673-3200
Address	Telephone
Upon filing.	
Adopted on	
07/01/2013 thru 12/28/2013	
Effective dates	

RULE CAPTION

Temporary rule regarding the implementation of End Tidal CO2 monitoring effective December 29, 2013.

Not more than 15 words

RULEMAKING ACTION

ADOPT: 818-026-0140

AMEND:

SUSPEND:

Stat. Auth.: ORS 679.250 (7)

Other Auth.: None

Stats. Implemented: ORS 670.250

RULE SUMMARY

The Board is creating Temporary Rule 810-026-0140 Implementation Date of End Tidal CO2 Monitoring.

STATEMENT OF NEED AND JUSTIFICATION

Rulemaking - Anesthesia Monitoring

In the Matter of

Guidelines for the Use of Sedation and General Anesthesia by Dentists adopted by the October 2007 American Dental Association House of Delegates, Guidelines for Teaching Pain Control and Sedation to Dentists and Dental Students adopted by the October 2012 American Dental Association House of Delegates, Policy Statement: The Use of Sedation and General Anesthesia by Dentists adopted by the October 2007 American Dental Association House of Delegates.

Documents Relied Upon, and where they are available

The Board needs to adopt Temporary Rule OAR 818-026-0140 Implementation Date of End Tidal CO2 Monitoring in order to give Licensees additional time to be in compliance with OAR 818-026-0060 (7)(a) Moderate Sedation Permit (7)(a), OAR 818-026-0065 (7)(a) Deep Sedation and OAR 818-026-0070(7)(a) General Anesthesia.

Need for the Temporary Rule(s)

The Board has learned that there is currently a backlog in availability of equipment that Licensees need to purchase in order to be in compliance with OAR 818-026-0060 (7)(a), OAR 818-026-0065 (7)(a) and OAR 818-026-0070(7)(a).

Justification of Temporary Rules

DRAFT

DIVISION 26

818-026-0140

Implementation Date of End Tidal CO2 Monitoring

The requirement to have patients continuously monitored with End-tidal CO2 monitors as provided in OAR 818-026-0060(7)(a), OAR 818-026-0065(7)(a) and OAR 818-026-0070 (7)(a) will become effective on January 1, 2014.

Stat. Auth.: ORS 679.250(7)

Stats. Implemented: ORS 679.250