

**OAR 333-064-0~~10~~100, Exhibit
C**

Table 1 – Accuracy requirements in percent recovery (%R) for laboratory control samples (LCS), initial calibration verification (ICV), and continuing calibration verification (CCV).:-

Test	Analyte	LCS Limits (%R)	ICV and CCV Limits (%R)
Pesticides in accordance with OAR 333-007-0 40 400	Abamectin	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Acephate	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Acequinocyl	40 40.0 - 160 60.0	<u>70.0 – 130.0</u>
	Acetamiprid	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Aldicarb	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Azoxystrobin	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Bifenazate	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Bifenthrin	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Boscalid	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Carbaryl	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Carbofuran	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Chlorantraniliprole	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Chlorfenapyr	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Chlorpyrifos	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Clofentezine	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Cyfluthrin	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Cypermethrin	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Daminozide	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	DDVP (Dichlorvos)	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Diazinon	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Dimethoate	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Ethoprophos	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>

	Etofenprox	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Etoxazole	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Fenoxycarb	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Fenpyroximate	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Fipronil	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Flonicamid	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Fludioxonil	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Hexythiazox	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Imazalil	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Imidacloprid	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Kresoxim-methyl	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Malathion	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Metalaxyl	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
Test	Analyte	LCS Limits (%R)	ICV and CCV Limits (%R)
Pesticides in accordance with OAR 333-007-040400 continued	<u>Methiocarb</u>	<u>60.0 - 120.0</u>	<u>70.0 – 130.0</u>
	<u>Methomyl</u>	<u>60.0 - 120.0</u>	<u>70.0 – 130.0</u>
	Methyl parathion	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	MGK-264	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Myclobutanil	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Naled	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Oxamyl	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Paclobutrazol	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Permethrins	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Phosmet	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Piperonyl butoxide	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Prallethrin	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Propiconazole	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Propoxur	60 60.0 -	<u>70.0 – 130.0</u>

		120 120.0	
	Pyrethrins	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Pyridaben	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Spinosad	50 50.0 - 150 50.0	<u>70.0 – 130.0</u>
	Spiromesifen	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Spirotetramat	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Spiroxamine	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Tebuconazole	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Thiacloprid	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Thiamethoxam	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
	Trifloxystrobin	60 60.0 - 120 120.0	<u>70.0 – 130.0</u>
Residual Solvents in accordance with OAR 333-007-04 10 <u>10</u>	1,4-Dioxane	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	2-Butanol	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	2-Ethoxyethanol	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	2-Propanol (IPA)	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Acetone	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Acetonitrile	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Benzene	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Butanes	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Cumene (aka isopropylbenzene)	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Cyclohexane	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Dichloromethane	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Ethyl acetate	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>
	Ethyl benzene	60 60.0 - 120 120.0	<u>60.0 - 120.0</u>

Test	Analyte	LC <u>SLCS</u> -Limits (%R)	<u>ICV and CCV</u> Limits (%R)
Residual Solvents in accordance with OAR 333-007-041010 continued	<u>Ethyl ether</u>	<u>60.0 - 120.0</u>	<u>60.0 - 120.0</u>
	<u>Ethylene glycol</u>	<u>60.0 - 120.0</u>	<u>60.0 - 120.0</u>
	<u>Ethylene Oxide</u>	<u>60.0 - 120.0</u>	<u>60.0 - 120.0</u>
	Heptane	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Hexanes	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Isopropyl acetate	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Methanol	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Pentanes	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Propane	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Tetrahydrofuran	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Toluene	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
	Xylenes	60 <u>60.0</u> - 120 <u>120.0</u>	<u>60.0 - 120.0</u>
Adult use cannabinoids and CBD in accordance with OAR 333-007-0430	delta-9 THC	90 <u>90.0</u> - 110 <u>110.0</u>	<u>90.0 - 110.0</u>
	delta-9 THCA	90 <u>90.0</u> - 110 <u>110.0</u>	<u>90.0 - 110.0</u>
	delta-8 THC	90 <u>90.0</u> - 110 <u>110.0</u>	<u>90.0 - 110.0</u>
	CBD	90 <u>90.0</u> - 110 <u>110.0</u>	<u>90.0 - 110.0</u>
	CBDA	90 <u>90.0</u> - 110 <u>110.0</u>	<u>90.0 - 110.0</u>
Heavy Metals in accordance with OAR 333-007-0415	Arsenic	80 <u>80.0</u> - 115 <u>115.0</u>	<u>90.0 - 110.0</u>
	Cadmium	80 <u>80.0</u> - 115 <u>115.0</u>	<u>90.0 - 110.0</u>
	Lead	80 <u>80.0</u> - 115 <u>115.0</u>	<u>90.0 - 110.0</u>
	Mercury	80 <u>80.0</u> - 115 <u>115.0</u>	<u>90.0 - 110.0</u>
Mycotoxins in accordance with OAR 333-007-0425	Aflatoxin B1	60 <u>60.0</u> - 120 <u>120.0</u>	<u>70.0 – 130.0</u>
	Aflatoxin B2	60 <u>60.0</u> - 120 <u>120.0</u>	<u>70.0 – 130.0</u>
	Aflatoxin G1	60 <u>60.0</u> - 120 <u>120.0</u>	<u>70.0 – 130.0</u>
	Aflatoxin G2	60 <u>60.0</u> - 120 <u>120.0</u>	<u>70.0 – 130.0</u>
	Ochratoxin A	60 <u>60.0</u> - 120 <u>120.0</u>	<u>70.0 – 130.0</u>
<u>Water Activity in accordance with OAR 333-007-0420</u>	<u>Water Activity</u>		<u>95.0 – 105.0</u>