

OAR 333-064-0140, Exhibit D

Table 1 – Accuracy requirements in percent recovery (%R) for initial calibration verification (ICV), continuing calibration verification (CCV), and laboratory control samples (LCS).

Test	Analyte	<u>ICV, CCV, and</u> <u>LCS Limits (%R)</u>
Potency in accordance with OAR 333-333-7040	Psilocybin	70. <u>0</u> – 130. <u>0</u>
	Psilocin	70. <u>0</u> – 130. <u>0</u>
Solvent testing in accordance with OAR 333-333-7050	Methanol	60. <u>0</u> – 120. <u>0</u>
	Acetic acid	60. <u>0</u> – 120. <u>0</u>