

GENERAL TECHNICAL CT REQUIREMENTS IN VETERINARY CLINICS

- **Machine Physics Report**

An Oregon approved CT medical physicist shall complete an initial performance evaluation before first use and annually thereafter. A performance evaluation also needs to be completed after any change in the equipment which might cause a change in radiation output or image quality by a CT medical physicist.

- **Shielding Study**

A shielding survey will also need to be completed by a CT medical physicist. Any change in placement of the CT machine changes the shielding requirements. If your machine is on wheels, you need to be very specific about what you plan to do with it.

GENERAL TECHNICAL FLUOROSCOPY REQUIREMENTS IN VETERINARY CLINICS

- **Machine Physics Report**

An Oregon approved Medical Physicist must complete an initial performance evaluation before first use and annually thereafter. A performance evaluation also needs to be completed after any change in the equipment which might cause a change in radiation output or image quality by a medical physicist.

- **Shielding Survey**

Will need to be completed by an Oregon approved Medical Physicist. Any change in placement of the fluoroscopy machine changes the shielding requirements. If your machine is on wheels you need to be very specific about what you plan to do with it.