

Approved Handheld units for use in Oregon

Manufacturer	Model	Dental	Medical	Comments
Aribex/Kavo/Dexis	Nomad Nomad Pro 2	✓ ✓		
Dexcowin	DX3000/Maxray Cocoon 60/1.7	✓ ✓		
Digidoc	Xray2go / Minix-s	✓		
Mobilex	T100	✓		
Remex	T100	✓		You might see Remidi but we register it under Remex T-100
Vatech	EzRay Air P: Vex P300	✓		
Genoray	Zen-PX2	✓		also known as Port-XII but discontinued as of 05/18/2021 per Vendor
	Zen-PX4	✓		also known as Port-XIV
Carestream	CS2400P	✓		
Ecotron	DT-703	✓		Approved 12/05/2022. Tube Model is OX/70 and Generator Model is EMB-DT703 in case they submit the wrong serial numbers. Never use the Cradle info. Reapproved 07/13/2023
Flight Dental Sys	XR-3800	✓		Also might see X - Vision listed on documentation but should have XR-3800 on the machine itself. Approved 3/2/23
Runyes	RAY98(P)	✓		Approved 9/14/2023 Manuel Version 4)
Digimed	DHX-70H	✓		Approved 4/19/2024 for the system with the shield that has no holes. The wafer-like shield that has holes is not approved so be sure to check to make sure it has the proper shield.

current as of 04/19/2024