Off-site Vaccination Clinic
Temperature Log for Refrigerator- Fahrenheit

Date of Clinic:_____ **Clinic Hours:** _____

Facility Name: _____

Off-site Clinic Location: _____

Monitor temperatures closely!

1. Write your initials below in “Staff Initials,” and note the time in “Exact Time.”
2. Record the min/max tempatures of the vaccine after transport to off-site clinic.
3. Record vaccine tempatures every hour during the clinic by putting an X in the row that corresponds with the storage unit temperature.
4. At the end of the day record the min/max temperatures reached.
5. Record min/max temps again after returning vaccine to main clinic.
6. If any out-of-range temps, see instructions to the right.

Take action if temp is out of range—too warm (above 46ºF) or too cold (below 46ºF).

1. Relocate the vaccines to proper conditions as quickly as possible. Label the vaccines "Do Not Use". Do not discard vaccines unless directed to by the Oregon VFC Program
2. Record the out-of-range temps in the “Action” area on the bottom of the log.
3. Call the Oregon VFC Helpdesk at 971-673-4832 to determine vaccine viability and next steps.
4. Document the action taken on the “Vaccine Storage Troubleshooting Record"

Clinic Hour	Transport	Hour 1	Hour 2	Hour 3	Hour 4	Hour 5	Hour 6	Hour 7	Hour 8	Hour 9	Hour 10	Hour 11	Hour 12	End of Day	Transport
Staff Initials															
Exact Time															
Danger! Temperatures above 46°F are too Warm! Write any out-of-range temps on the lines below and call VFC Helpdesk (971-673-4832) immediately!															
Acceptable Temperatures 46°F	Max:													Max:	Max:
45°F															
44°F															
43°F															
42°F															
Aim for 41° 41°F															
40°F															
39°F															
38°F															
37°F															
36°F	Min:													Min:	Min:
Danger! Temperatures below 36°F are too cold! Write any out-of-range temps on the lines below and call VFC Helpdesk (971-673-4832) immediately!															
Action — Write any out-of-range temps (above 46°F or below 36°F) here:															

'